

**CMPT Water Microbiology Survey, H212
Final Drinking Water Results for HPC
(shipped July 12, 2021)**

Samples		Group Mean/Median/MU%	
		Total Count (cfu/ml)	MU%
1	<i>Escherichia coli</i>	186/191	18.5
2	<i>Escherichia coli</i>	196/197	15.7
3	no organisms present	*0/0	*0.0
4	<i>Enterobacter species</i>	*86/88	*28.6

*outliers were not included in group mean/median/MU calculations