

**CMPT Water Microbiology Survey, R212,
Preliminary Recreational Water Results
(shipped August 23, 2021)**

Method	Group Mean/Median/MU%		
	R212-1: <i>Pseudomonas aeruginosa</i>	R212-2: <i>Escherichia coli</i>	R212-3: <i>Enterococcus species</i>
Membrane Filtration cfu/100 ml	*127/123/18.4	112/110/16.4	112/111/18.6
Enzyme Substrate/MPN MPN/100ml	*132.9/140.0	*107.9/109.0	*109.3/97.6

*outliers were not included in group mean/median/MU calculations

MU% - Measurement Uncertainty percentage calculated for Membrane Filtration method only